



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that LabCorp Central Laboratory, using Roche Cobas c503 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: <b>Cobas c503</b>	Calibrator Lot: <b>559759, 559759</b>	Application: <b>(HbA1c/Hb)*91.5+2.15</b>
Reagent Lot: <b>645103, 611695, 657543</b>	Calibrator Assigned Values: <b>0.153, 0.348, 0.660, 0.903, 1.24, 1.75 µmol/L HbA1c; 1.45, 7.78 µmol/L Hb</b>	

Date of Certification: January 1, 2023

Certification Expires: January 1, 2024

*Daniel Sacks*

NGSP Steering Committee Chair

*Randie R Little PhD.*

NGSP Network Coordinator

*Sharon Amulof*

SRL director/ supervisor